

B&B SCUBA INC.

NA

Test Site

BAUER

ID #'s

Client ID: NA

Agent ID: PRES96701

PRESSURE SYSTEMS, INC.

Program Information

Program #:

Frequency: Single

Kit #: **275983**

Tracking Information

Date Received: 06/24/13

Issue ID: 246009

CGA E (G-7.1, 2011)


Analytes	Test #1				Range
	06/27/13				
Oxygen (%)	21.2				20-22
Carbon Dioxide (ppm)	350.87				1000
Carbon Monoxide (ppm)	<1				10
T.H.C. (ppm)	<1				25
Water Vapor (ppm)	19.81				67
Dew Point (F)	-67				-50
Oil/Part. (mg/m3)	<1				5
Unknowns	ND				NA
Odor	ND				None

Spectra #: 246009.SPA

Results: Sample Meets Standard

S
P
E
C
T
R
A
L

R
E
A
D
O
U
T



COPYRIGHT © 1999 Lawrence Factor, Inc.