



Certificate of Analysis

University of Hawaii at Manoa

is in compliance with the air/gas quality portion of the specification:

Oxygen Compatible Air-2003 (I)

as analyzed and reported on this certificate for the sample described under section "sample & report information"



Analytical Test Methods:		Media Sampled:	Estimate of Uncertainty:
Gases & Vapors:	CAT-A-01 Gas Chromatography/Mass Spectrometry	Source Bottle: 786387	The average estimate of uncertainty at standard specification limits for 10 compounds normally reported is $\pm 3.24\%$, at a 95% confidence interval ($k=2$). For more detailed uncertainty information, contact Trace Analytics, LLC.
Oil & Particulate:	CAT-A-03 Analytical Gravimetry	Ambient Bottle: N/A	
Particle Size:	CAT-A-04 Optical Microscopy	Source Filter: 8311	
Pressure Dew Point:	CAT-A-07 Gas Detector Tube	Detector Tube: Draeger 5-a/P	

Maria Sandoval
Maria Sandoval, Laboratory Director

Results of Test: PASS

Sample & Report Information

From: Trace Analytics, LLC 15768 Hamilton Pool Road Austin, Texas 78738 To: University of Hawaii at Manoa EHSO- Diving Safety Program c/o Michael Pamatat 2040 East West Road Honolulu, HI 96822	Analytes	Source Results (1)	Ambient Results	Specification Allowable Limits	Sampled For	University of Hawaii at Manoa
	Oxygen, Volume %	20.8	N/A	20-22	Sampled By	Mike Pamatat
	Nitrogen, Volume %	78.2	N/A	N/A	Sampled On	1/15/2021
	Argon, Volume %	1.0	N/A	N/A	Received On	1/27/2021
	Nitrogen Plus Argon, Volume %	79.2	N/A	N/A	Analyzed On	2/1/2021
	Carbon Monoxide (CO), ppmv	<0.5	N/A	2	Sampled From	Compressor & Stored Air
	Carbon Dioxide (CO2), ppmv	171	N/A	1000	Make	Hypress
	Water Content (H2O), ppmv/Dewpoint, °F (DT)	14 / -72	N/A	N/A / N/A (W)	Model	HP-6000-NA4-E3
	TVHC (including CH4), ppmv	5.2	N/A	25	Serial No.	1A070302
	Methane (CH4) ppmv	<0.7	N/A	N/A	Cylinder(s)	Bank 1-7
	TVHC (excluding CH4), ppmv	5.2	N/A	N/A		
	Oil (condensed) & Particulate, mg/m3	<0.02	N/A	0.1		
	Odor (provided by customer)	None/Slight	N/A	None/Slight	Hours	966.4
Other	N/A	N/A	N/A	Sample Phase	Routine	
Other	N/A	N/A	N/A	Customer Comments		
Other	N/A	N/A	N/A	Report Number	21-02868	
				Customer ID	3267	
				Date Reported	2/1/2021	
				Frequency	Quarterly	

(1) Results apply to the sample as received from the customer. Information supplied by the customer can affect the validity of results.

(DT) Water content Pass/Fail was determined by water vapor detector tube analysis.

(I) This specification for oxygen compatible air is taken from ANSI/CGA G-7.1 Grade E as modified by International Association of Nitrox and Technical Divers (IANTD) in their document Blending Standards, 2003.

(W) Dew point is expressed in °F at one atmosphere pressure absolute.




15768 Hamilton Pool Road
 Austin, Texas 78738
 800-AIR-1024 or 512-263-0000 • Fax: 512-263-0002
 E-mail: ServiceTeam@AirCheckLab.com

Routine AirCheck™ Datasheet

*IF RUSH REQUESTED, CALL WITH TRACKING NUMBER

DS-BA-01

1 Contact Information IMPORTANT: FILL OUT COMPLETELY, CAREFULLY PRINT, AND RETURN A DATA SHEET WITH EACH SAMPLE SET.

Customer ID **3267**  Customer Name **University of Hawaii at Manoa** Country **USA**
 Primary Mr. Mike Pamatat E-mail **uhdiveop@hawaii.edu** Phone **(808) 956-7179**
 Alternate **Mr. Dave Pence** E-mail **uhdiveop@hawaii.edu** Phone **(808) 956-6420**

Please fill in the circle to the left if you'd like the AirCheck Report sent to the person below (fill in information).

Additional E-mail

2 Rush Analysis Request

RUSH REQUESTED, \$125 extra, Initial Here By initialing, I am authorizing Same Day* Analysis & Reporting for an add'l \$125 per sample. **CALL CUST. SERVICE @ EXT 3 TO SCHEDULE ~ Samples must arrive by 10:30 a.m.** Contact Us for Holiday Scheduling

3 Purchase Order Information (if applicable) **5 Customer Comments (use back if needed)**

If a purchase order number is required by your company, please attach it to this data sheet and write the PO Number in the space provided here.

PO Number

4 System Information Wrong Wrong Wrong Correc **6 Sampled By and Sample Date**

System ID

Sampled For

Testing Schedule 45 Days Monthly Startup
 90 Days Other Verification
 120 Days Quarterly Weekly
 Annual Random Sample
 Bimonthly Semi-Annual

Air Spec

Make

Model

Serial No

Cylinder

Other ID

Pressure High Pressure (1,000-6,000 psi)
 Low Pressure (less than 1,000 psi)

Air used for SCBA Airline Respirator
 SCUBA Other

Purification Molecular Sieve/Desiccant No Purification
 Refrigerated Dryer Unknown
 No Dryer

Sampled From Compressor Source Other
 Stored Air Outlet Not Provided
 Comp. & Storage Breather Box

Sample Phase Before Filter Change After Filter Change Routine

Comp. Hours

Lowest Temp °F °C

(Lowest temp, low pressure breathing air may be exposed to during the year)

I attest that all information provided on this datasheet is truthful and accurate to the best of my knowledge. Submittal of this air sample authorizes Trace Analytics, LLC to provide services

SIGNATURE _____ PRINT Name (Person taking the test sample) _____
 Date Sample Taken / /
MONTH DAY YEAR

7 Sample Information

Is this sample a Retest taken within 30 days of a failed test? Yes No

A Source Bottle, Filter, and Data Sheet **MUST BE RETURNED** for a complete analysis.

Filter Number (red or green label) 6 or 7 digits

Flowrate (liters per minute)

Sample Time (minimum of 10 min.)

Detector Tube (OMIT data if sampling media does not include Detector Tube)

Tube Reading (0 - 200) DT Minutes Sampled

Source Bottle Number (blue label) 6 or 7 digits

Ambient Bottle Number (white label) 6 or 7 digits

Odor is **REQUIRED**. MARK ONLY ONE. None Detectable Slight Pronounced
 It's determined by sniffing the air from the side port of the Bottle Holder.

PLEASE NOTE - NO EXCEPTIONS

Sample Shelf Life: Once a sample is taken, it must be received by our laboratory within 60 days.

Component Media Shelf Life: Sampling media must be used or returned for free replacement within 2 years of shipment date. See expiration date on return box.

For TRACE Use Only

Receiving I.D.

Receiver's Initials

We Do One Thing – Test Compressed Air

www.AirCheckLab.com

Sampling Notes for Water Vapor Detector Tube

- 1: Break BOTH tips of detector tube before inserting. Arrow on tube points away from Fitting. 50 LPM for 10 minutes.
- 2: The DT is filled with a chemical reagent that reacts to the presence of water by changing color from yellow to a grayish/reddish brown. (ignore gray color)

A minimum of 500 liters of air is required for sampling. If you are unable to achieve 50 LPM, adjust sampling time using the following formula:

$$\frac{500 \text{ LITERS}}{\text{FLOWRATE}} = \text{SAMPLE TIME, MIN}$$

Detector Tube Results are based on a 10 minute sample at 50 LPM.

Sampling for longer or shorter time periods will provide different results than shown on chart. Identify the farthest reddish-brown color change in the tube (ignore gray coloration). Locate where the DT reading and Flowrate (LPM) intersect to determine approximate result in F*. If results do not meet your air specification limits, take corrective actions, and repeat sample. If both samples are returned at the same time, the 2nd sample will be at no charge. For troubleshooting tips, go to: <https://rb.gy/gixijy>



Det. Tube Reading, mg/m ³	2.5	5	10	20	30	40	50	60	70	80	90	100	125	175	200
Flowrate Reading (LPM) 60	-93	-84	-75	-66	-60	-56	-52	-49	-47	-45	-43	-42	-38	-33	-31
55	-92	-83	-74	-65	-58	-54	-51	-48	-45	-44	-42	-40	-36	-31	-29
50	-90	-81	-72	-62	-56	-52	-49	-46	-44	-42	-40	-38	-34	-29	-27
45	-88	-79	-70	-60	-54	-50	-47	-44	-41	-39	-38	-36	-32	-26	-24
40	-86	-77	-68	-58	-52	-47	-44	-41	-39	-36	-35	-33	-29	-23	-21
PASS							FAIL								

Above area marked "Pass" is for high pressure air used for SCBA, with a -65°F limit per CGA Grade L/NFPA
 Visit the AirCheck Academy for complete range of flow rates and further details.

If detector tube reading is higher than required, see the following checklist or visit the Aircheck Academy: Breathing Air. www.AirCheckLab.com

Purification / Depressurized filters	High ambient air temperatures (above 70°F) affect the operating life of the cartridge. Chemicals used in purification filters begin to degrade as soon as they are installed. Is it time to change the filters?
Manual/auto drain or priority valve	If not working properly can be source for excess water and reduce filter life.
Remote fill or hose reel	Long lengths (>10 ft) of hose are notorious for accumulating and retaining water. A short 1-2 minute purge WILL NOT be sufficient. It is best to take the sample from a short fill hose (5-10 ft) or directly from containment fill station. - View our resource
Recent hydrostat	Bottles must be properly dried after hydrostat and should be immediately pressurized with dry air.
Valves left open	Ambient air can easily have 10,000 - 50,000 ppm of water. Purge sufficiently to remove water accumulated from ambient air.
Sample taken from storage	Take sample from compressor to identify if compressor is producing dry air. If yes, storage banks may contain excess water. Drain and refill with dry air. This may require 2-3 fills to drive off water from inside cylinders. You can request extra detector tubes.
Detector tube cracked	Only the tips of the tube should be broken. If a crack runs down the main body of the tube, results will not be dependable.
Tube fitting wet	If multiple samples are taken consecutively, excess water may pool inside the fitting. Dry fitting between uses.
Other	Keep in mind that 1 milliliter (which is about 20 drops from an eyedropper) in a 1.7 cubic ft cylinder at 4500 psig would be 90 ppm of water vapor. It doesn't take much to fail.

KIT UPGRADES AVAILABLE

We have redesigned our detector tube assembly and tip breaker for ease of use. Available for K901c, K901s, and K901n AirCheck Kits. Please contact our customer service team to order your upgrade today: (512) 263-0000 ext. 3.

Included in Kit Upgrade (U902):

- Brass adapter with wire mesh
- Detector tube holder
- Tip Breaker

